

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-G5 RYE GRASS

| | | | |
|---------|----------|---------|---------|
| Code | 88000-G5 | | |
| Lot No. | 985 | 2024-10 | 2023-11 |

REAGENTS

| Label | Description | Lot. | Exp. | NO |
|-------|----------------------------------|------|---------|-----|
| D9901 | 88000-G5 RYE GRASS | 985 | 2024-10 | 2x6 |
| B9900 | CALIBRATORE 88000-G5 RYE GRASS | 985 | 2024-10 | 1 |
| C9900 | CONTROL SERUM 88000-G5 RYE GRASS | 985 | 2024-10 | 1 |

CONTROLS

| | Found results | Expected results |
|----------------------|---------------|------------------|
| Calibrator (mOD) | 267 | ≥ 100 |
| Control Serum (KU/L) | 9,5 | 7,5-15,7 |

CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 6 | 3 | 1 | 9 | 8 | 5 | 6 | 5 | 1 | 8 | 6 | 2 | 2 | 4 | 8 | 7 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature

631985651862248700
Rye-grass
CALIBRATOR 0.250 mL [IVD]
LOT 985
 2024-10 2°C 8°C
DIESSE Monteriggioni Italy CE
B9900

CONTROL SERUM

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 6 | 3 | 1 | 9 | 8 | 5 | 6 | 3 | 0 | 5 | 5 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature

631985630550000001
Rye-grass
CONTROL 0.450 mL [IVD]
LOT 985
 2024-10 2°C 8°C
DIESSE Monteriggioni Italy CE
C9900

*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 08/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.