

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** 88000-E1 CAT EPITHELIUM

Code	88000-E1		
Lot No.	985	2024-10	2023-11

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D1191	88000-E1 CAT EPITHELIUM	985	2024-10	2x6
B1190	CALIBRATORE 88000-E1 CAT EPITHELIUM	985	2024-10	1
C1190	CONTROL SERUM 88000-E1 CAT EPITHELIUM	985	2024-10	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	267	≥ 100
Control Serum (KU/L)	9,5	7,5-15,7

**CALIBRATOR**

6	0	6	9	8	5	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

606985651862246700  
Cat epithelium  
CALIBRATOR 0.250 mL IVD  
LOT 985  
2024-10 2°C-8°C B1190  
DIESSE Monteriggioni Italy CE

**CONTROL SERUM**

6	0	6	9	8	5	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

606985630550000001  
Cat epithelium  
CONTROL+ 0.450 mL IVD  
LOT 985  
2024-10 2°C-8°C C1190  
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 08/11/2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.