



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-D1 DERMATOPHAGOIDES PTERONYSSINUS

Code	88000-D1		
Lot No.	985	2024-10	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D9791	88000-D1 DERMATOPHAGOIDES PTERONYSSINUS	985	2024-10	2x6
B9790	CALIBRATORE 88000-D1 DERMATOPHAGOIDES PTERONYSSINUS	985	2024-10	1
C9790	CONTROL SERUM 88000-D1 DERMATOPHAGOIDES PTERONYSSINUS	985	2024-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	267	≥ 100
Control Serum (KU/L)	9,5	7,5-15,7

CALIBRATOR

6	0	2	9	8	5	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

602985651862248700
Dermatophagoides
pteronyssinus
CALIBRATOR 0.250 mL [IVD]
LOT 985
2024-10 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

6	0	2	9	8	5	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

602985630550000001
Dermatophagoides
pteronyssinus
CONTROL 0.450 mL [IVD]
LOT 985
2024-10 2°C-8°C
DIESSSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 08/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.