



DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME tTg-A - CHORUS

Code	86058		
Lot No.	978	2024-10	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D3431	tTg-A	978	2024-10	6X6
B3431	CALIBRATORE tTg-A	978	2024-10	1
C3431	CONTROL SERUM tTg-A	978	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	633	≥ 100
Control Serum (AU/mL)	25,7	21,7-49,1

339978621930513700  
tTg-A  
CALIBRATOR 0.178 mL IVD  
LOT 978  
2024-10  
DIESSE Monteriggioni Italy CE

CALIBRATOR

3 3 9 9 7 8 6 2 1 9 3 0 5 1 3 7 0 0

Label checked with barcode reader YES  Operator Signature

CONTROL SERUM

3 3 9 9 7 8 6 1 4 2 6 0 0 0 0 0 0 1

Label checked with barcode reader YES  Operator Signature

33997861426000001  
tTg-A  
CONTROL+ 0.428 mL IVD  
LOT 978  
2024-10  
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 09-11-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.