



DIESSSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-M6 ALTERNARIA ALTERNATA

Code	88000-M6		
Lot No.	985	2024-10	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D1181	88000-M6 ALTERNARIA ALTERNATA	985	2024-10	2X6
B1180	CALIBRATORE 88000-M6 ALTERNARIA ALTERNATA	985	2024-10	1
C1180	CONTROL SERUM 88000-M6 ALTERNARIA ALTERNATA	985	2024-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	267	≥ 100
Control Serum (KU/L)	9,5	7,5-15,7

CALIBRATOR

6	0	7	9	8	5	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

607985651862248700
Alternaria alternata
CALIBRATOR 0.250 mL IVD
LOT 985
2024-10 2°C
DIESSSE Monteriggioni Italy CE
B1180

CONTROL SERUM

6	0	7	9	8	5	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

607985630550000001
Alternaria alternata
CONTROL SERUM 0.450 mL IVD
LOT 985
2024-10 2°C
DIESSSE Monteriggioni Italy CE
C1180

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... FB

DATE 08/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.