



QUALITY CONTROL CERTIFICATE

KIT NAME LKM-1 CHORUS

Code	86080		
Lot No.	987	2024-11	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D3641	LKM-1	987	2024-11	6X6
B3641	CALIBRATORE LKM-1	987	2024-11	1
C3641	CONTROL SERUM LKM-1	987	2024-11	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	607	≥ 100
Control Serum AU/mL	81,1	53,6-130,8

CALIBRATOR

7	0	3	9	8	7	6	4	1	9	3	3	3	5	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature HB

703987641933355900
LKM-1
CALIBRATOR 0.175 mL [IVD]
LOT 987
2024-11 2°C-8°C B3641
DIESSE Monteriggioni Italy CE

CONTROL SERUM

7	0	3	9	8	7	6	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature HB

703987620370000001
LKM-1
CONTROL 0.425 mL [IVD]
LOT 987
2024-11 2°C-8°C C3641
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents Blue

DATE 14/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.