

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** a-TPO - CHORUS

|                |         |         |         |
|----------------|---------|---------|---------|
| <b>Code</b>    | 8607412 |         |         |
| <b>Lot No.</b> | 981     | 2024-10 | 2023-10 |

**REAGENTS**

| Label | Description         | Lot. | Exp.    | NO  |
|-------|---------------------|------|---------|-----|
| D3591 | a-TPO               | 981  | 2024-10 | 2x6 |
| B3591 | CALIBRATORE a-TPO   | 981  | 2024-10 | 1   |
| C3591 | CONTROL SERUM a-TPO | 981  | 2024-10 | 1   |

**CONTROLS**

|                       | Found results | Expected results |
|-----------------------|---------------|------------------|
| Calibrator (mOD)      | 391           | ≥ 100            |
| Control Serum (IU/mL) | 350,1         | 264,3-593,8      |

**CALIBRATOR**

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 3 | 0 | 1 | 9 | 8 | 1 | 6 | 5 | 2 | 3 | 5 | 1 | 2 | 4 | 6 | 7 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES  Operator Signature

301981652351246700  
a-TPO  
CALIBRATOR 0.176 mL [IVD]  
LOT 981  
 2024-10 2°C-8°C  
DIESSE Monteriggioni Italy CE

**CONTROL SERUM**

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 3 | 0 | 1 | 9 | 8 | 1 | 6 | 2 | 9 | 5 | 4 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES  Operator Signature

301981629540000001  
a-TPO  
CONTROL SERUM 0.425 mL [IVD]  
LOT 981  
 2024-10 2°C-8°C  
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 16-11-2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.