



**QUALITY CONTROL CERTIFICATE**

**KIT NAME** TOXOPLASMA IgM - CHORUS

Code	81041		
Lot No.	978	2024-10	2023-11

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D2281	TOXOPLASMA IgM	978	2024-10	6X6
B2281	CALIBRATORE TOXOPLASMA IgM	978	2024-11	1
C2281	CONTROL SERUM TOXOPLASMA IgM	978	2024-11	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	380	≥ 100
Control Serum (INDEX)	3,1	1,9-3,5
Negative Control (INDEX)	0,1	<0,7

**CALIBRATOR**

1	3	1	9	7	8	6	0	0	0	0	0	3	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *Green*

13197860000370000  
Toxoplasma IgM  
CALIBRATOR 0.175 mL IVD  
LOT 978  
2024-11 2°C-8°C  
DIESSE Monteriggioni Italy CE 0123

**CONTROL SERUM**

1	3	1	9	7	8	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *Green*

131978618350000001  
Toxoplasma IgM  
CONTROL+ 0.425 mL IVD  
LOT 978  
2024-11 2°C-8°C  
DIESSE Monteriggioni Italy CE 0123

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Green*

DATE *15-11-2023*

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.