



QUALITY CONTROL CERTIFICATE

KIT NAME CARDIOLIPIN-G - CHORUS

Code	86046		
Lot No.	981	2024-10	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D3311	CARDIOLIPIN-G	981	2024-10	6X6
B3311	CALIBRATORE CARDIOLIPIN-G	981	2024-10	1
C3311	CONTROL SERUM CARDIOLIPIN-G	981	2024-10	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	540	≥ 100
Control Serum (GPLU/mL)	35,8	26,5-58,0

CALIBRATOR

3	3	5	9	8	1	6	3	2	3	3	8	3	5	5	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature RB



CONTROL SERUM

3	3	5	9	8	1	6	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature RB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 20/11/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.