



QUALITY CONTROL CERTIFICATE

KIT NAME CARDIOLIPIN-M - CHORUS

Code	86048		
Lot No.	980	2024-10	2023-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D3331	CARDIOLIPIN-M	980	2024-10	6X6
B3331	CALIBRATORE CARDIOLIPIN-M	980	2024-10	1
C3331	CONTROL SERUM CARDIOLIPIN-M	980	2024-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	616	≥ 100
Control Serum (MPLU/mL)	48,6	32,5-90,0

CALIBRATOR

3	3	6	9	8	0	6	7	1	9	3	1	3	5	3	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

336980671931353900
Cardiolipin-M
CALIBRATOR 0.176 mL IVD
LOT 980
2024-10
DIESSE Monteriggioni Italy CE

CONTROL SERUM

3	3	6	9	8	0	6	2	0	3	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

336980620370000001
Cardiolipin-M
CONTROL+ 0.426 mL IVD
LOT 980
2024-10
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents... *FB*

DATE 22.11.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.