



DIESE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

QUALITY CONTROL CERTIFICATE

KIT NAME CARDIOLIPIN-G - CHORUS

Code 86046  
Lot No. 978 2024-10 2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D3311	CARDIOLIPIN-G	978	2024-10	6X6
B3311	CALIBRATORE CARDIOLIPIN-G	978	2024-11	1
C3311	CONTROL SERUM CARDIOLIPIN-G	978	2024-11	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	545	≥ 100
Control Serum (GPLU/mL)	27,6	23,2-52,5

CALIBRATOR

3 3 5 9 7 8 6 3 2 1 3 3 4 5 4 9 0 0

Label checked with barcode reader YES  Operator Signature *Brew*

335978632133454900  
Cardiolipin-G  
CALIBRATOR 0.175 mL IVD  
LOT 978  
2024-11 2°C-8°C  
DIESE Monteriggioni Italy CE

CONTROL SERUM

3 3 5 9 7 8 6 1 6 3 0 0 0 0 0 0 0 1

Label checked with barcode reader YES  Operator Signature *Brew*

335978616300000001  
Cardiolipin-G  
CONTROL+ 0.425 mL IVD  
LOT 978  
2024-11 2°C-8°C  
DIESE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Brew*

DATE 24.11.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.