

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** Vitamin D Total- CHORUS

Code	86900		
Lot No.	978	2024-10	2023-11

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D9702	25OH Vitamin D Total	978	2024-10	6x6
B9703	CALIBRATOR 25OH Vitamin D Total	978	2025-11	1
C9703	CONTROL SERUM 25OH Vitamin D Total	978	2025-11	4

**CONTROLS**

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	523	≥ 100
CONTROL SERUM (ng/ml)	33,8	30,3-38,5

**CALIBRATOR**

8	3	5	9	7	8	6	1	3	4	3	7	5	8	5	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *SS*

835978613437585600  
25OH Vitamin D Total  
CALIBRATOR 0.300 mL [IVD]  
LOT 978  
2025-11 2°C-8°C  
DIESSE Monteriggioni Italy CE 0123

**CONTROL SERUM**

8	3	5	9	7	8	6	1	2	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *SS*

835978612230000001  
25OH Vitamin D Total  
CONTROL+ 0.300 mL [IVD]  
LOT 978  
2025-11 2°C-8°C  
DIESSE Monteriggioni Italy CE 0123

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *SS*

DATE 29.11.2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.