



DIESSSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** CARDIOLIPIN-G - CHORUS

Code	86046
Lot No.	988
	🕒 2024-11
	📄 2023-11

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D3311	CARDIOLIPIN-G	988	2024-11	6X6
B3311	CALIBRATORE CARDIOLIPIN-G	988	2024-11	1
C3311	CONTROL SERUM CARDIOLIPIN-G	988	2024-11	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	361	≥ 100
Control Serum (GPLU/mL)	41,4	34,3-85,3

**CALIBRATOR**

3	3	5	9	8	8	6	4	2	6	4	0	2	5	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *[Signature]*

335988642640255900  
Cardiolipin-G  
CALIBRATOR 0.175 mL IVD  
LOT 988  
🕒 2024-11 2°C-8°C  
DIESSSE Monteriggioni Italy CE

**CONTROL SERUM**

3	3	5	9	8	8	6	2	8	5	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *[Signature]*

335988628520000001  
Cardiolipin-G  
CONTROL+ 0.425 mL IVD  
LOT 988  
🕒 2024-11 2°C-8°C  
DIESSSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents: *[Signature]*

DATE 23.11.2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.