



QUALITY CONTROL CERTIFICATE

KIT NAME Q FEVER PHASE II IgG - CHORUS

Code	81171		
Lot No.	970	2024-09	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D2080	Q FEVER PHASE II IgG	970	2024-09	6x6
B2080	CALIBRATORE Q FEVER PHASE II IgG	970	2024-09	1
C2080	CONTROL SERUM Q FEVER PHASE II IgG	970	2024-09	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	345	≥ 100
Control Serum (Index)	2,1	1,9-3,5

CALIBRATOR

9	4	4	9	7	0	6	0	0	0	0	0	3	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

94497060000370000
Q Fever phase II IgG
CALIBRATOR 0,175 mL IVD
LOT 970
2024-09 2°C-8°C
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

9	4	4	9	7	0	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

94497061935000001
Q Fever phase II IgG
CONTROL+ 0,425 mL IVD
LOT 970
2024-09 2°C-8°C
DIESSSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 29.11.23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.