

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F49 APPLE

Code	88000-F49		
Lot No.	985	2024-10	2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D1061	88000-F49 APPLE	985	2024-10	2X6
B1060	CALIBRATORE 88000-F49 APPLE	985	2024-10	1
C1060	CONTROL SERUM 88000-F49 APPLE	985	2024-10	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	267	≥ 100
Control Serum (KU/L)	9,5	7,5-15,7

CALIBRATOR

6	6	0	9	8	5	6	5	1	8	6	2	2	4	8	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

660985651862248700
Apple
CALIBRATOR 0.250 mL IVD
LOT 985
 2024-10 2°C-8°C
DIESE Monteriggioni Italy CE
B1060

CONTROL SERUM

6	6	0	9	8	5	6	3	0	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature

660985630550000001
Apple
CONTROL+ 0.450 mL IVD
LOT 985
 2024-10 2°C-8°C
DIESE Monteriggioni Italy CE
C1060

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 04.12.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.