



QUALITY CONTROL CERTIFICATE

KIT NAME BORRELIA IgM - CHORUS

Code	81047		
Lot No.	988	2024-11	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D2310	BORRELIA IgM	988	2024-11	6X6
B2310	CALIBRATORE BORRELIA IgM	988	2025-02	1
C2310	CONTROL BORRELIA IgM	988	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	430	≥ 100
Control (Index)	3,8	2,6-4,8

CALIBRATOR

9	0	5	9	8	8	6	0	0	0	0	0	2	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES Operator Signature



CONTROL SERUM

9	0	5	9	8	8	6	2	6	4	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 01/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.