



QUALITY CONTROL CERTIFICATE

KIT NAME ANA-8 - CHORUS

Code	86010		
Lot No.	988	2024-11	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D3111	ANA-8	988	2024-11	6X6
B3111	CALIBRATORE ANA-8	988	2024-11	1
C3111	CONTROL SERUM ANA-8	988	2024-11	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	450	≥ 100
Control Serum (index)	2,5	1,8-3,3

CALIBRATOR

1	0	0	9	8	8	6	0	0	0	0	0	4	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *MB*

10098860000400000  
ANA-8  
CALIBRATOR 0.175 mL IVD  
LOT 988  
2024-11 2°C-8°C B3111  
DIESSE Monteriggioni Italy CE

CONTROL SERUM

1	0	0	9	8	8	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *MB*

100988618330000001  
ANA-8  
CONTROL 0.425 mL IVD  
LOT 988  
2024-11 2°C-8°C B3111  
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *07.12.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.