

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione  
20144 Milano (Italy) Via A. Solari 19 sc. 6

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** ENZYWELL HERPES SIMPLEX 1 IgM

Code	91028		
LOT.NO	241	2025-02	2023-11

**REAGENTS**

CODE	LABEL	DESCRIPTION	LOT	EXP	NO
PF93000	007200	MICROPLATE	672	2025-04	1
PF93925	007190	CUT OFF CONTROL	219	2025-05	1
PF93900/HSV	C5020	NEGATIVE CONTROL	805	2025-05	1
PF93923	007220	POSITIVE CONTROL	207	2025-05	1
PF91721	007230	ANTIGEN	221	2026-09	6
PF93611	009660	DILUENT 2	231	2025-03	1
PF93603	1035	WASHING BUFFER 10x	464	2025-04	1
PF93602	1032/RISKO	STOP SOLUTION	218-5	2025-06	1
PF93619	1034	SUBSTRATE-HS	380	2025-03	1
PF30055	007180	CONJUGATE	226	2025-02	1

**CONTROLS**

	FOUND RESULTS	EXPECTED RESULTS
NEGATIVE CONTROL	0.197	< 0,6 Cut Off
CUT OFF CONTROL	0.520	≥ 0.200
POSITIVE CONTROL	1.782	> 1,5 Cut Off
CV %	7%	≤ 15%

*Evaluatin of the method according to the package inset in the attached document  
This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED BY: RESPONSIBLE HEAD, Quality Control Reagents

DATE: 15/11/23

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.

IO 09- 005-V



CODE 91028  
CODE 91186

## QUALITY CONTROL CERTIFICATE

Annex: Method Evaluation

NAME KIT HERPES SIMPLEX 1 IgM  
LOT. 241  
DATE OF TEST 14/11/2023

METHOD: ACCORDING TO THE PACKAGE INSERT

1. SOLID PHASE	FOUND	EXPECTED
A) UNIFORMITY BETWEEN PLATES		
1st PLATE	OD 450 nm 1.445 CV% 16%	< 20%
2nd PLATE	OD 450 nm 1.639 CV% 4%	< 20%
B) UNIFORMITY WITHIN PLATE		
	OD 450 nm 1.785 CV% 12%	< 15%
2. BULK CONTROL date: 09/11/2023		
A) Average OD of Cut Off Control	OD 450 nm 0.577	≥ 0.200
B) Index of Negative and Positive Sera	accepted	According to specification
C) Sensitivity with working calibrators	accepted	According to specification
3. FINAL CONTROL date:14/11/2023		
A) Blank	OD 450 nm 0.051	≤ 0.100
B) Average OD of Negative Control	OD 450 nm 0.197	< 0,6 Cut Off
C) Average OD of Cut Off Control	OD 450 nm 0.520	≥ 0.200
D) Average OD of Positive Control	OD 450 nm 1.782	> 1,5 Cut Off
E) HOMOGENEITY OF CUT OFF	No. 6 OD 450 nm 0.589 CV% 7%	≤ 15%
F) PRECISION OF FILLING PROCESS		

SATISFACTORY

UNSATISFACTORY

FINAL RESULT: ACCEPTED

Date: 14.11.2023

Signed:   
Released: Q.C.Manager