

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F20 ALMOND

Code	88000-F20		
Lot No.	985	2024-10	2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D1051	88000-F20 ALMOND	985	2024-10	2x6
B1050	CALIBRATORE 88000-F20 ALMOND	985	2025-02	1
C1050	CONTROL SERUM 88000-F20 ALMOND	985	2025-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	314	≥ 100
Control Serum (KU/L)	8,2	5,9-16,1

CALIBRATOR

6	7	3	9	8	5	6	5	2	0	5	1	2	8	9	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature 88



CONTROL SERUM

6	7	3	9	8	5	6	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature 88



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... geli

DATE 13/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.