



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F25 TOMATO

Code	88000-F25		
Lot No.	985	2024-10	2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D9851	88000-F25 TOMATO	985	2024-10	2x6
B9850	CALIBRATORE 88000-F25 TOMATO	985	2025-02	1
C9850	CONTROL SERUM 88000-F25 TOMATO	985	2025-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	314	≥ 100
Control Serum (KU/L)	8,2	5,9-16,1

CALIBRATOR

6	1	3	9	8	5	6	5	2	0	5	1	2	8	9	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *SB*

613985652051289600  
Tomato  
CALIBRATOR 0.280 mL IVD  
LOT 985  
2025-02 2°C -8°C  
DIESSE Monteriggioni Italy CE

CONTROL SERUM

6	1	3	9	8	5	6	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *SB*

613985625470000001  
Tomato  
CONTROL 0.450 mL IVD  
LOT 985  
2025-02 2°C -8°C  
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Alvisei*

DATE 13/12/2023

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**  
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.