



DIESSSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-G3 COCKSFOOT

Code	88000-G3		
Lot No.	985	2024-10	2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D9891	88000-G3 COCKSFOOT	985	2024-10	2X6
B9890	CALIBRATORE 88000-G3 COCKSFOOT	985	2025-02	1
C9890	CONTROL SERUM 88000-G3 COCKSFOOT	985	2025-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	314	≥ 100
Control Serum (KU/L)	8,2	5,9-16,1

CALIBRATOR

6	4	9	9	8	5	6	5	2	0	5	1	2	8	9	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *SB*



CONTROL SERUM

6	4	9	9	8	5	6	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *SB*



*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Alvina*

DATE 13/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.