

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-D2 DERMATOPHAGOIDES FARINAE

Code	88000-D2		
Lot No.	985	2024-10	2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D9801	88000-D2 DERMATOPHAGOIDES FARINAE	985	2024-10	2X6
B9800	CALIBRATORE 88000-D2 DERMATOPHAGOIDES FARINAE	985	2025-02	1
C9800	CONTROL SERUM 88000-D2 DERMATOPHAGOIDES FARINAE	985	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	314	≥ 100
Control Serum (KU/L)	8,2	5,9-16,1

CALIBRATOR

6	0	3	9	8	5	6	5	2	0	5	1	2	8	9	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature SB

603985652051289600
Dermatophagoides
farinae
CALIBRATOR 0.250 mL IVD
LOT 985
2026-02 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

6	0	3	9	8	5	6	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature SB

603985625470000001
Dermatophagoides
farinae
CONTROL+ 0.450 mL IVD
LOT 985
2026-02 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. SB

DATE 10.12.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!
This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.