



DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME Vitamin D Total- CHORUS

Code	86900		
Lot No.	979	2024-10	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D9702	25OH Vitamin D Total	979	2024-10	6x6
B9703	CALIBRATOR 25OH Vitamin D Total	979	2025-11	1
C9703	CONTROL SERUM 25OH Vitamin D Total	979	2025-11	4

CONTROLS

	Found results	Expected results
Calibrator (mOD)	571	≥ 100
CONTROL SERUM (ng/ml)	33,5	30,3-38,5

CALIBRATOR

8	3	5	9	7	9	6	1	3	4	3	7	5	8	5	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*



CONTROL SERUM

8	3	5	9	7	9	6	1	2	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 21/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.