

QUALITY CONTROL CERTIFICATE

KIT NAME ANTI-CCP - CHORUS

Code	86094		
Lot No.	988	2024-11	2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D8671	ANTI-CCP	988	2024-11	6x6
B8671	CALIBRATORE ANTI-CCP	988	2024-11	1
C8671	CONTROL SERUM ANTI-CCP	988	2024-11	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	340	≥ 100
Control Serum (AU/mL)	53,3	46,1-90,4

CALIBRATOR

2	5	6	9	8	8	6	3	2	8	5	4	2	7	7	2	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB



CONTROL SERUM

2	5	6	9	8	8	6	2	6	4	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. FB

DATE 28/12/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.