

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME LKM-1 CHORUS

Code 86080

Lot No. 997

2024-12

2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D3641	LKM-1	997	2024-12	6X6
B3641	CALIBRATORE LKM-1	997	2024-12	1
C3641	CONTROL SERUM LKM-1	997	2024-12	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	526	≥ 100
Control Serum (AU/mL)	46,8	22,8-60,7

CALIBRATOR

7 0 3 9 9 7 6 4 1 9 3 3 6 1 5 9 0 0


Label checked with barcode reader

YES



Operator Signature

[Signature]



CONTROL SERUM

7 0 3 9 9 7 6 1 2 2 2 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature

[Signature]



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE 22-12-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.