

QUALITY CONTROL CERTIFICATE

KIT NAME tTg-G - CHORUS

Code	86060		
Lot No.	982	2024-10	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D3451	tTg-G	982	2024-10	6X6
B3451	CALIBRATORE tTg-G	982	2024-10	1
C3451	CONTROL SERUM tTg-G	982	2024-10	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	588	≥ 100
Control Serum (AU/mL)	36,5	23,6-49,3

CALIBRATOR

3	4	0	9	8	2	6	2	1	7	3	1	5	1	4	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*



CONTROL SERUM

3	4	0	9	8	2	6	1	4	2	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *FB*

DATE *28/12/2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.