



QUALITY CONTROL CERTIFICATE

KIT NAME AMA-M2 - CHORUS

Code	86084		
Lot No.	988	2025-02	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D3651	AMA-M2	988	2025-02	6X6
B3651	CALIBRATORE AMA-M2	988	2025-02	1
C3651	CONTROL SERUM AMA-M2	988	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	400	≥ 100
Control Serum (AU/mL)	143,7	50,0-287,5

CALIBRATOR

7	0	5	9	8	8	7	3	2	5	3	2	2	9	4	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

705988732532294800
AMA-M2
CALIBRATOR 0.175 mL IVD
LOT 988
2025-02
DIESSE Monteriggioni Italy CE

CONTROL SERUM

7	0	5	9	8	8	7	2	5	4	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature MB

705988725470000001
AMA-M2
CONTROL 0.425 mL IVD
LOT 988
2025-02
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 08/01/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.