

QUALITY CONTROL CERTIFICATE

KIT NAME BETA 2-GLYCOPROTEIN-M - CHORUS

Code	86052		
Lot No.	989	2024-11	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D337I	BETA 2-GLYCOPROTEIN-M	989	2024-11	6X6
B337I	CALIBRATORE BETA 2-GLYCOPROTEIN-M	989	2024-11	1
C337I	CONTROL SERUM BETA 2-GLYCOPROTEIN-M	989	2024-11	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	370	≥ 100
Control Serum (AU/mL)	93,1	50,0-157,1

CALIBRATOR

2	0	5	9	8	9	6	3	2	4	3	9	2	9	6	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

205989632439296000
Beta-2-Glycoprotein-M
CALIBRATOR 0.175 mL IVD
LOT 989
2024-11
DIESSE Monteriggioni Italy CE

CONTROL SERUM

2	0	5	9	8	9	6	2	4	4	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

205989624450000001
Beta-2-Glycoprotein-M
CONTROL 0.425 mL IVD
LOT 989
2024-11
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents FB

DATE 08/10/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.