



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F13 PEANUT

Code	88000-F13		
Lot No.	996	2024-12	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D9941	88000-F13 PEANUT	996	2024-12	2x6
B9940	CALIBRATORE 88000-F13 PEANUT	996	2025-02	1
C9940	CONTROL SERUM 88000-F13 PEANUT	996	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	310	≥ 100
Control Serum (KU/L)	9,9	5,8-14,3

CALIBRATOR

6	1	7	9	9	6	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

617996652160268100
Peanut
CALIBRATOR 0.250 mL IVD
LOT 996
2025-02 2°C 8°C B9940
DIESSE Monteriggioni Italy CE

CONTROL SERUM

6	1	7	9	9	6	6	2	7	5	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *FB*

617996627500000001
Peanut
CONTROL+ 0.450 mL IVD
LOT 996
2025-02 2°C 8°C C9940
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *FB*

DATE 10/01/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.