



QUALITY CONTROL CERTIFICATE

KIT NAME ZIKA IgM CAPTURE -CHORUS

Code	81161		
Lot No.	996	2024-12	2023-12

REAGENTS

Label	Description	Lot.	Exp	No
D4290	Zika IgM Capture	996	2024-12	6X6
B4290	Calibratore Zika IgM Capture	996	2024-12	1
C4290	Control Zika IgM Capture	996	2024-12	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	400	> 100
Control (Index)	2,5	1,8-3,3

CALIBRATOR

9	8	7	9	9	6	6	0	0	0	0	0	4	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

98799660000400000  
Zika IgM Capture  
CALIBRATOR 0.480 mL IVD  
LOT 996  
2024-12  
DIESSE Monteriggioni Italy CE

CONTROL SERUM

9	8	7	9	9	6	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

987996618330000001  
Zika IgM Capture  
CONTROL 0.900 mL IVD  
LOT 996  
2024-12  
DIESSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 09/01/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.