

QUALITY CONTROL CERTIFICATE

KIT NAME DEAMIDATED GLIADIN-G - CHORUS

Code	86100		
Lot No.	989-A	2024-11	2023-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D8660	Deamidated Gliadin-G	989	2024-11	6X6
B8660	CALIBRATORE Deamidated Gliadin-G	989-1	2024-12	1
C8660	CONTROL SERUM Deamidated Gliadin-G	989-1	2024-12	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	620	≥ 100
Control Serum (AU/mL)	60,4	34,1-87,5

CALIBRATOR

2	5	5	9	8	9	6	1	2	1	3	3	4	3	5	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

255989612133435600
DEAMIDATED GLIADIN-G
CALIBRATOR 0.175 mL [IVD]
LOT 989-1
2024-12 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

2	5	5	9	8	9	6	1	6	3	0	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

255989616300000001
DEAMIDATED GLIADIN-G
CONTROL+ 0.425 mL [IVD]
LOT 989-1
2024-12 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... FB

DATE 15/01/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.