



QUALITY CONTROL CERTIFICATE

KIT NAME ANA-Screen - CHORUS

Code	86014		
Lot No.	996	2024-12	2023-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D3141	ANA-Screen	996	2024-12	6X6
B3141	CALIBRATORE ANA-Screen	996	2024-12	1
C3141	CONTROL SERUM ANA-Screen	996	2024-12	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	383	≥ 100
Control Serum (Ndx)	3,2	2,0-3,6

CALIBRATOR

1	1	5	9	9	6	6	0	0	0	0	0	3	6	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

11599660000360000  
ANA-Screen  
CALIBRATOR 0.178 mL IVD  
LOT 996  
2024-12 2°C-8°C  
DIESSE Monteriggioni Italy CE

CONTROL SERUM

1	1	5	9	9	6	6	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

115996620360000001  
ANA-Screen  
CONTROL+ 0.425 mL IVD  
LOT 996  
2024-12 2°C-8°C  
DIESSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents FB

DATE 16/01/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.