



DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME U1-70 RNP - CHORUS

Code	86088		
Lot No.	996	2024-12	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D3662	U1-70 RNP	996	2024-12	6X6
B3662	CALIBRATORE U1-70 RNP	996	2024-12	1
C3662	CONTROL SERUM U1-70 RNP	996	2024-12	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	518	≥ 100
Control Serum (AU/mL)	56,3	26,7-83,3

CALIBRATOR

1	0	4	9	9	6	6	2	1	9	2	8	5	3	3	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature HB

104996621928533900  
U1-70 RNP  
CALIBRATOR 0.175 mL IVD  
LOT 996  
 2024-12 2°C-8°C  
DIESSE Monteriggioni Italy CE

CONTROL SERUM

1	0	4	9	9	6	6	1	3	2	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature HB

104996613250000001  
U1-70 RNP  
CONTROL+ 0.425 mL IVD  
LOT 996  
 2024-12 2°C-8°C  
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE A.01.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.