

QUALITY CONTROL CERTIFICATE

KIT NAME 88000-D1 DERMATOPHAGOIDES PTERONYSSINUS

Code	88000-D1		
Lot No.	996	2024-12	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D9791	88000-D1 DERMATOPHAGOIDES PTERONYSSINUS	996	2024-12	2x6
B9790	CALIBRATORE 88000-D1 DERMATOPHAGOIDES PTERONYSSINUS	996	2025-02	1
C9790	CONTROL SERUM 88000-D1 DERMATOPHAGOIDES PTERONYSSINUS	996	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	310	≥ 100
Control Serum (KU/L)	9,9	5,8-14,3

CALIBRATOR

6	0	2	9	9	6	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

602996652160268100
Dermatophagoides pteronyssinus
CALIBRATOR 0.250 mL IVD
LOT 996
2025-02 2°C-8°C
DIESSÉ Monteriggioni Italy CE

CONTROL SERUM

6	0	2	9	9	6	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

602996627500000001
Dermatophagoides pteronyssinus
CONTROL+ 0.450 mL IVD
LOT 996
2025-02 2°C-8°C
DIESSÉ Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents FB

DATE 22.01.2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.