

QUALITY CONTROL CERTIFICATE

KIT NAME HELICOBACTER PYLORI IgG -CHORUS

Code	81060		
Lot No.	001	2025-01	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D2391	Helicobacter Pylori IgG	001	2025-01	6X6
B2392	CALIBRATORE Helicobacter Pylori IgG	001	2025-01	1
C2391	CONTROL SERUM Helicobacter Pylori IgG	001	2025-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	410	≥ 100
Control Serum (UA/mL)	101,6	75,0-188,9
Negative Control (UA/mL)	<5,0	< 5

CALIBRATOR

0	0	8	0	0	1	6	1	2	0	3	2	2	7	5	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

008001612032275000
Helicobacter pylori IgG
CALIBRATOR 0.425 mL IVD
LOT 001
2025-01
2°C-8°C
DIESSE Monteriggioni Italy CE
B2392

CONTROL SERUM

0	0	8	0	0	1	6	2	6	4	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

008001626480000001
Helicobacter pylori IgG
CONTROL+ 0.425 mL IVD
LOT 001
2025-01
2°C-8°C
DIESSE Monteriggioni Italy CE
C2391

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. Agnes

DATE 24/01/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.