



DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME SANITIZING SOLUTION

Code	83604		
Lot No.	074	2027-01	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
B2563	SANITIZING SOLUTION	074	2027-01	1

CONTROLS

	Found results	Expected results
Density	1.061	1.060-1.100
pH	13,4	12-14

*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 29.01.24