

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** 88000-M6 ALTERNARIA ALTERNATA

Code	88000-M6		
Lot No.	001	2025-01	2024-02

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D1181	88000-M6 ALTERNARIA ALTERNATA	001	2025-01	2X6
B1180	CALIBRATORE 88000-M6 ALTERNARIA ALTERNATA	001	2025-02	1
C1180	CONTROL SERUM 88000-M6 ALTERNARIA ALTERNATA	001	2025-02	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

**CALIBRATOR**

6	0	7	0	0	1	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

607001652160268100  
Alternaria alternata

CALIBRATOR 0,250 mL IVD

LOT 001

2025-02 2°C B1180

**DIESSSE** Monteriggioni Italy

**CONTROL SERUM**

6	0	7	0	0	1	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

607001627500000001  
Alternaria alternata

CONTROL+ 0,450 mL IVD

LOT 001

2025-02 2°C C1180

**DIESSSE** Monteriggioni Italy

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *FB*

DATE *01.02.2024*

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.