



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-W21 WALL PELLITORY W21

Code	88000-W21		
Lot No.	001	2025-01	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D9961	88000-W21 WALL PELLITORY W21	001	2025-01	2X6
B9960	CALIBRATORE 88000-W21 WALL PELLITORY W21	001	2025-02	1
C9960	CONTROL SERUM 88000-W21 WALL PELLITORY W21	001	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

CALIBRATOR

6	1	6	0	0	1	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

616001652160268100  
Wall pellitory W21  
CALIBRATOR 0.250 mL [IVD]  
LOT 001  
2025-02 2°C B9960  
DIESSSE Monteriggioni Italy

CONTROL SERUM

6	1	6	0	0	1	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature *FB*

616001627500000001  
Wall pellitory W21  
CONTROL+ 0.450 mL [IVD]  
LOT 001  
2025-02 2°C C9960  
DIESSSE Monteriggioni Italy

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *Aluis*

DATE *d. 02. 2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.