



QUALITY CONTROL CERTIFICATE

KIT NAME **tTg-G - CHORUS**

Code	86060		
Lot No.	001	2025-01	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D3451	tTg-G	001	2025-01	6x6
B3451	CALIBRATORE tTg-G	001	2025-01	1
C3451	CONTROL SERUM tTg-G	001	2025-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	350	≥ 100
Control Serum (AU/mL)	42,0	31,9-70,0

CALIBRATOR

3	4	0	0	0	1	6	1	1	9	3	5	3	6	3	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

340001611935363900
tTg-G
CALIBRATOR 0.175 mL IVD
LOT 001
2025-01 2°C-8°C
DIESSE Monteriggioni Italy CE

CONTROL SERUM

3	4	0	0	0	1	6	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature *MB*

340001620360000001
tTg-G
CONTROL+ 0.425 mL IVD
LOT 001
2025-01 2°C-8°C
DIESSE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE *02/02/2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.