

QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA PNEUMONIAE IgG - CHORUS

Code	81034		
Lot No.	001	2025-01	2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D2251	MYCOPLASMA PNEUMONIAE IgG	001	2025-01	6x6
B2252	CALIBRATORE MYCOPLASMA PNEUMONIAE IgG	001	2025-01	1
C2251	CONTROL SERUM MYCOPLASMA PNEUMONIAE IgG	001	2025-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	440	≥ 100
Control Serum (UA/mL)	38,7	24,4-58,0
Negative Control (UA/mL)	<10	<10

CALIBRATOR

0	6	6	0	0	1	6	8	1	2	2	1	4	9	4	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature sb

066001681221494100
Mycoplasma pneumoniae IgG
CALIBRATOR 0.425 mL IVD
LOT 001
2025-01 2°C-8°C B2252
DIESSA Monteriggioni Italy CE

CONTROL SERUM

0	6	6	0	0	1	6	1	4	2	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature sb

066001614270000001
Mycoplasma pneumoniae IgG
CONTROL 0.425 mL IVD
LOT 001
2025-01 2°C-8°C C2251
DIESSA Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents sb

DATE 02.02.24

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.