

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TREPONEMA IgG -CHORUS

Code 81050

Lot No. 996

2024-12

2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D2321	TREPONEMA IgG	996	2024-12	6X6
B2322	CALIBRATORE TREPONEMA IgG	996	2025-01	1
C2321	CONTROL SERUM TREPONEMA IgG	996	2025-01	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	300	≥ 100
Control Serum (INDEX)	3,3	2,1-3,9
Negative Control (INDEX)	0,0	<0,7


CALIBRATOR

0	1	2	9	9	6	6	0	0	0	0	0	0	3	4	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *SS*



CONTROL SERUM

0	1	2	9	9	6	6	2	1	3	9	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *SS*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *SS*

DATE 02.02.24

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.