



QUALITY CONTROL CERTIFICATE

KIT NAME CYTOMEGALOVIRUS IgG AVIDITY CHORUS


Code 81092
Lot No. 001 2025-01 2024-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D2491	CYTOMEGALOVIRUS IgG AVIDITY	001	2025-01	2x6
C2491	CONTROL SERUM HIGH CMV AVIDITY	001	2025-01	1
G2491	SAMPLE DILUENT	001	2025-04	1

CONTROLS

	Found results	Expected results
Control Serum (%)	72,5%	50-95%



CALIBRATOR

Label checked with barcode reader YES Operator Signature

CONTROL SERUM

0 8 4 0 0 1 6 5 0 9 5 0 0 0 0 0 0 0 1

Label checked with barcode reader YES Operator Signature *SB*

084001650950000001
Cytomegalovirus IgG Avidity
CONTROL HIGH 0.3 mL IVD
LOT 001
2025-01 2°C-8°C
DIESSE Monteriggioni Italy CE 0123

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Bus*

DATE *02.02.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.