



**QUALITY CONTROL CERTIFICATE**

**KIT NAME 88000-T210 PRIVET**

Code	88000-T210		
Lot No.	001	2025-01	2024-02

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D6861	88000-T210 PRIVET	001	2025-01	2x6
B6860	CALIBRATORE 88000-T210 PRIVET	001	2025-02	1
C6860	CONTROL SERUM 88000-T210 PRIVET	001	2025-02	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

**CALIBRATOR**

7	9	5	0	0	1	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES  Operator Signature FB



**CONTROL SERUM**

7	9	5	0	0	1	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB



*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... [Signature]

DATE 08/02/2024

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.