



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F85 CELERY

|         |           |         |         |  |
|---------|-----------|---------|---------|--|
| Code    | 88000-F85 |         |         |  |
| Lot No. | 001       | 2024-10 | 2024-02 |  |

REAGENTS

| Label | Description                    | Lot. | Exp.    | NO  |
|-------|--------------------------------|------|---------|-----|
| D0961 | 88000-F85 CELERY               | 001  | 2024-10 | 2X6 |
| B0960 | CALIBRATORE 88000-F85 CELERY   | 001  | 2025-02 | 1   |
| C0960 | CONTROL SERUM 88000-F85 CELERY | 001  | 2025-02 | 1   |

CONTROLS

|                      | Found results | Expected results |
|----------------------|---------------|------------------|
| Calibrator (mOD)     | 284           | ≥ 100            |
| Control Serum (KU/L) | 8,2           | 5,8-14,3         |

CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 6 | 8 | 3 | 0 | 0 | 1 | 5 | 5 | 2 | 1 | 6 | 0 | 2 | 6 | 8 | 1 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES  Operator Signature FB

683001552160268100  
Celery  
CALIBRATOR 0.250 mL IVD  
LOT 001  
2025-02 2°C-8°C  
DIESSSE Monteriggioni Italy CE

CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 6 | 8 | 3 | 0 | 0 | 1 | 5 | 2 | 7 | 5 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES  Operator Signature FB

683001527500000001  
Celery  
CONTROL 0.450 mL IVD  
LOT 001  
2025-02 2°C-8°C  
DIESSSE Monteriggioni Italy CE

*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. [Signature]

DATE 08/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.