



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-T11 PLANE

Code	88000-T11		
Lot No.	007	2024-11	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D1121	88000-T11 PLANE	007	2024-11	2X6
B1120	CALIBRATORE 88000-T11 PLANE	007	2025-02	1
C1120	CONTROL SERUM 88000-T11 PLANE	007	2025-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

CALIBRATOR

6	7	8	0	0	7	5	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature



CONTROL SERUM

6	7	8	0	0	7	5	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature



*This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 08/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.