



DIESSSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

**QUALITY CONTROL CERTIFICATE**

**KIT NAME** VARICELLA IgM CHORUS

Code	81079		
Lot No.	997	2025-03	2023-12

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D2481	VARICELLA IgM	997	2025-03	6x6
B2481	CALIBRATORE VARICELLA IgM	997	2025-03	1
C2481	CONTROL SERUM VARICELLA IgM	997	2025-03	1

**CONTROLS**

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	540	≥ 100
Control Serum (Index)	2,5	1,8-3,3
Negative Control (Index)	0,0	< 0,7

**CALIBRATOR**

1	3	4	9	9	7	7	0	0	0	0	0	0	4	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature MB

134997700000400000  
Varicella IgM  
CALIBRATOR 0.175 mL IVD  
LOT 997  
2025-03  
DIESSSE Monteriggioni Italy CE

**CONTROL SERUM**

1	3	4	9	9	7	7	1	8	3	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature MB

134997718330000001  
Varicella IgM  
CONTROL 0.425 mL IVD  
LOT 997  
2025-03  
DIESSSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents [Signature]

DATE 09/02/2024

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.