



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-P1 ASCARIS

Code	88000 - P1		
Lot No.	001	2025-01	2024-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D6851	88000-P1 ASCARIS	001	2025-01	2X6
B6850	CALIBRATORE 88000-P1 ASCARIS	001	2025-02	1
C6850	CONTROL SERUM 88000-P1 ASCARIS	001	2025-02	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

CALIBRATOR

7	9	3	0	0	1	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

793001652160268100
Ascaris
CALIBRATOR 0.250 mL [IVD]
LOT 001
 2025-02 2024-02
DIESE Monteriggioni Italy CE

CONTROL SERUM

7	9	3	0	0	1	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES Operator Signature FB

793001627500000001
Ascaris
CONTROL 0.450 mL [IVD]
LOT 001
 2025-02 2024-02
DIESE Monteriggioni Italy CE

This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... FB

DATE 05/02/2024

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.