

**QUALITY CONTROL CERTIFICATE**

**KIT NAME 88000-W204 SUNFLOWER**

Code	88000-W204		
Lot No.	002	2025-01	2024-02

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D6881	88000-W204 SUNFLOWER	002	2025-01	2X6
B6880	CALIBRATORE 88000-W204 SUNFLOWER	002	2025-02	1
C6880	CONTROL SERUM 88000-W204 SUNFLOWER	002	2025-02	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

**CALIBRATOR**

7	9	7	0	0	2	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

797002652160268100  
Sunflower  
CALIBRATOR 0.250 mL IVD  
LOT 002  
2025-02 2°C -8°C  
DIESSE Monteriggioni Italy CE

**CONTROL SERUM**

7	9	7	0	0	2	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature

797002627500000001  
Sunflower  
CONTROL+ 0.450 mL IVD  
LOT 002  
2025-02 2°C -8°C  
DIESSE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 15.02.2024

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.