



QUALITY CONTROL CERTIFICATE

KIT NAME 88000-W2 WESTERN RAGWEED

| | | | |
|---------|----------|---------|---------|
| Code | 88000-W2 | | |
| Lot No. | 002 | 2025-01 | 2024-02 |

REAGENTS

| Label | Description | Lot. | Exp. | NO |
|-------|--|------|---------|-----|
| D3151 | 88000-W2 WESTERN RAGWEED | 002 | 2025-01 | 2X6 |
| B3150 | CALIBRATORE 88000-W2 WESTERN RAGWEED | 002 | 2025-02 | 1 |
| C3150 | CONTROL SERUM 88000-W2 WESTERN RAGWEED | 002 | 2025-02 | 1 |


CONTROLS

| | Found results | Expected results |
|----------------------|---------------|------------------|
| Calibrator (mOD) | 284 | ≥ 100 |
| Control Serum (KU/L) | 8,2 | 5,8-14,3 |

CALIBRATOR

| | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 6 | 8 | 6 | 0 | 0 | 2 | 6 | 5 | 2 | 1 | 6 | 0 | 2 | 6 | 8 | 1 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader YES Operator Signature *SB*



CONTROL SERUM

| | | | | | | | | | | | | | | | | | | |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 6 | 8 | 6 | 0 | 0 | 2 | 6 | 2 | 7 | 5 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES Operator Signature *SB*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *SB*

DATE *15.02.2024*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.