

**QUALITY CONTROL CERTIFICATE**

**KIT NAME 88000-W12 GOLDEN ROD**

Code	88000-W12		
Lot No.	002	2025-01	2024-02

**REAGENTS**

Label	Description	Lot.	Exp.	NO
D6871	88000-W12 GOLDEN ROD	002	2025-01	2X6
B6870	CALIBRATORE 88000-W12 GOLDEN ROD	002	2025-02	1
C6870	CONTROL SERUM 88000-W12 GOLDEN ROD	002	2025-02	1

**CONTROLS**

	Found results	Expected results
Calibrator (mOD)	284	≥ 100
Control Serum (KU/L)	8,2	5,8-14,3

796002652160268100  
Golden rod

CALIBRATOR 0.250 mL IVD  
LOT 002  
2025-02 2°C-8°C  
DIESE Monteriggioni Italy CE

**CALIBRATOR**

7	9	6	0	0	2	6	5	2	1	6	0	2	6	8	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

**CONTROL SERUM**

7	9	6	0	0	2	6	2	7	5	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES  Operator Signature FB

796002627500000001  
Golden rod

CONTROL+ 0.450 mL IVD  
LOT 002  
2025-02 2°C-8°C  
DIESE Monteriggioni Italy CE

**This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.**

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... FB

DATE 19/02/2024

**WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!**

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.